ePM 10M/12M/15M PATIENT MONITORS

Performance Specifications for Ambulatory Surgery Center and Clinic

Display

Type:	Capacitive touchscreen, supports multi-touch operation.
Size and Resolution:	
ePM 10M:	10.1" – 1280 x 800 pixels
ePM 12M:	12.1" – 1280 x 800 pixels
ePM 15M:	15.6" – 1366 x 768 pixels
Waveforms:	
ePM 10M:	Up to 8 waveforms
ePM 12M:	Up to 10 waveforms
ePM 15M:	Up to 12 waveforms

ECG (3, 5 lead)

Leads:	I, II, III, aVR, aVL, aVF, V
Sweep Speed:	6.25 mm/s, 12.5 mm/s, 25 mm/s, 50 mm/s
Gain Selection:	x 0.125, x 0.25, x 0.5, x 1, x 2, x 4, auto
Input Signal Range:	±8 mV (p-p)
Electrode Offset Potential Tolerance:	±500 mV
Bandwidth	
Diagnostic Mode:	0.05 to 150 Hz
Monitor Mode:	0.5 to 40 Hz
Surgical Mode:	1 to 20 Hz
ST Mode:	0.05 to 40 Hz
Defibrillator Protection:	Withstand 5000 V (360J) voltage in isolation against and defibrillation
Recovery Time:	≤5 s
CMRR	
Diagnostic:	>90 dB
Monitor, Surgical, ST Mode:	>105 dB (with notch filter on)
ESU Recovery Time:	≤10 s

Heart Rate Meter

Measurement Range:	
Adult:	15 to 300 bpm
Pediatric/Neonate:	15 to 350 bpm
Accuracy:	±1 bpm or ±1%, whichever is greater
Resolution:	1 bpm
Tall T-wave Rejection:	When the test is performed based on Clause 201.12.1.101.17 of ANSI/AAMI/IEC 60601-2-27, the heart rate calculation is not affected for QRS of 1 mV amplitude and 100 ms duration, T-wave duration of 180 ms and amplitude lower than 1.2 mV, and QT interval of 350 ms.

Pace Pulse

Pace:	Pace pulses meeting the following conditions are labelled with a PACE marker:
Amplitude:	±2 to ±700 mV
Width:	0.1 to 2 ms
Rise Time:	10 to 100 µs (without overshoot)
Pace Pulse Rejection:	When tested in accordance with the ANSI/AAMI/ IEC 60601-2-27: 201.12.1.101.13, the heart rate meter rejects all pulses meeting the following conditions.
Amplitude:	±2 to ±700 mV
Width:	0.1 to 2 ms
Rise Time:	10 to 100 µs (without overshoot)

Arrhythmia Analysis

Patient:	Adult/Pediatric
Monitored Arrhythmias:	Asystole, V-Fib/V-Tach, V-Tach, Vent Brady, Extreme Tachy, Extreme Brady, Vent Rhythm, PVCs/min, Pauses/min, Couplet, Bigeminy, Trigeminy, R on T, Run PVCs, PVC, Tachy, Brady, Missed Beats, Pacer Not Pacing, Pacer Not Capture, Multiform PVC, Nonsus V-Tach, Pause, Irr Rhythm, A-Fib



ST Segment Analysis (optional)

Patient:	Adult/Pediatric/Neonate
Range:	-2.0 to +2.0 mV (RTI)
Accuracy:	-0.8 to +0.8 mV: ±0.02 mV or ±10%, whichever is greater Beyond this range: not specified
Resolution:	0.01 mV
ST Adjust Scale:	40 ms after J points, 60 ms after J point (default), 80 ms after J point and 60/80 ms after J point (80 ms when HR 120 bpm or less; 60 ms when HR above 120 bpm)
ISO Adjust Scale:	4-200 ms before R-Wave (default: 80 ms), Step: 4 ms
J Point Adjust Scale:	4-200 ms after R-Wave (default: 48 ms), Step: 4 ms

QT Analysis (optional)

Patient:	Adult/Pediatric/Neonate	
Parameters:	QT, QTc, ΔQTc	
QTc Formula:	Bazett, Fridericia, Framingham, or Hodges	
Range:		
QT/QTc:	200 to 800 ms	
QT-HR:	Adult:	15 to 150 bpm
	Pediatric/Neonate:	15 to 180 bpm
QT Accuracy:	±30 ms	
Resolution:	QT 4 ms; QTc 1 ms	

Respiration

Lead:	I or II, auto
RR Range:	0 to 120 rpm (Adult) 0 to 150 rpm (Pediatric/Neonate)
RR Accuracy:	± 2 rpm (7 to 150 rpm) or ± 2 %, whichever is greater 0 to 6 rpm: not specified
RR Resolution:	1 rpm
Sweep Speed:	3 mm/s, 6.25 mm/s, 12.5 mm/s, 25 mm/s, 50 mm/s
Apnea Alarm Delay:	10 s, 15 s, 20 s, 25 s, 30 s, 35 s, 40 s

Pulse Oximetry

With Masimo SET® SpO_2

Range:	1 to 100%
Resolution:	1%
Accuracy:	±2% (70 to 100%, Adult/Pediatric, non- motion) ±3% (70 to 100%, Neonate, non-motion) ±3% (70 to 100%, motion) 1 to 69% unspecified
Pulse Rate Range:	25 to 240 bpm
Pulse Rate Accuracy:	±3 bpm (non- motion) ±5 bpm (motion)
Refresh Rate	≤ 1s

mindray NORTH AMERICA

ePM 10M/12M/15M PATIENT MONITORS

Performance Specifications for Ambulatory Surgery Center and Clinic

Pulse Oximetry (continued)

With Nellcor SpO_2 (optional)

Range:	0 to 100%
Resolution:	1%
Accuracy:	±2% (70 to 100%, Adult/Pediatric)
	±3% (70 to 100%, Neonate)
	Unspecified (0 to 69%)
Pulse Rate Range:	20 to 300 bpm
Pulse Rate Accuracy:	±3 bpm (20 - 250 bpm)
	Unspecified (251 - 300 bpm)
Refresh Rate:	≤ 1 s

PR

Range:	25 to 350 bpm (from IBP) 30 to 300 bpm (from NIBP)
Accuracy:	±1 bpm or ±1 %, whichever is greater (from IBP)
	± 3 bpm or ±3 %, whichever is greater (from NIBP)

Temperature

Meet standard of ISO 80601-2-56.

Method:	Thermal resistance
Channels:	2 channels
Range:	0 to 50 °C (32 to 122 °F)
Accuracy:	± 0.1 °C or ± 0.2 °F (without probe)
Resolution:	0.1 °C

Non-Invasive Blood Pressure

Meet standards of ISO 80601-2-30.

Method:	Oscillometry	
Operation Mode:	Manual, Auto, STAT, Sequence	
Max Measurement Time:	Adult/Pediatric:	180 s
	Neonate:	90 s
Systolic Range:	Adult:	25 to 290 mmHg
	Pediatric:	25 to 240 mmHg
	Neonate:	25 to 140 mmHg
Diastolic Range:	Adult:	10 to 250 mmHg
	Pediatric:	10 to 200 mmHg
	Neonate:	10 to 115 mmHg
Mean Range:	Adult:	15 to 260 mmHg
	Pediatric:	15 to 215 mmHg
	Neonate:	15 to 125 mmHg
Accuracy:	Max mean error:	± 5 mmHg
	Max standard deviation:	8 mmHg
Resolution:	1 mmHg	
Assisting Venous Puncture:	Yes	

Invasive Blood Pressure (optional)

Meet standards of IEC 60601-2-34: 2011.

Technique:	Direct invasive measurement
Measurement Range:	-50 to 300 mmHg
Resolution:	1 mmHg
Accuracy:	±2% or ±1 mmHg, whichever is greater (excluding sensor error)
Refreshing Rate:	≤ 1 s
Recovery Time:	<10 s (after defibrillation)

Sidestream CO_2 (optional)

Measurement Range:	
$etCO_2$:	0 to 150 mmHg
CO2 Accuracy:	
0 to 40 mmHg:	±2 mmHg
41 to 76 mmHg:	±5% of reading
77 to 150 mmHg:	±10% of reading
Resolution:	
$etCO_2$:	1 mmHg
Sample Flow Rate:	
Adult/Pediatric:	120 ml/min
Neonate:	70 ml/min or 90 ml/min
Sample Flow Rate Tolerance:	±15 ml/min or ±15%, whichever is greater
Warm-up Time:	90 s (maximum), 20 s (typically)

Measured with a Neonatal watertrap and 2.5-meter Neonatal sampling line, or an Adult watertrap and a 2.5-meter Adult sampling line:

Rise Time:	
$etCO_2$:	≤250 ms @ 70 ml/min (Neonate watertrap)
	≤250 ms @ 90 ml/min (Neonate watertrap)
	≤300 ms @ 120 ml/min (Adult/Pediatric watertrap)
Response Time:	
$etCO_2$:	≤5.0 s @ 70 ml/min (Neonate watertrap)
	≤4.5 s @ 90 ml/min (Neonate watertrap)
	≤5.0 s @ 120 ml/min (Adult/Pediatric watertrap)
awRR Range:	0 to 150 rpm
awRR Accuracy:	
0 to 60 rpm:	±1 rpm
61 to 150 rpm:	±2 rpm
Apnea Alarm Delay:	10, 15, 20, 25, 30, 35, 40 s

Oridion Microstream CO_2 (optional)

Measurement Range:	0 to 99 mmHg
Resolution:	1 mmHg
Accuracy:	
0 to 38 mmHg:	±2 mmHg
39 to 99 mmHg:	±5% of the reading (0.08% increased in error for every 1 mmHg if the reading is more than 38 mmHg)
Sample Flow Rate:	50 -7.5 +15 ml/min
Start-up Time:	30 s (typical), 180 s (max)
awRR Range:	0 to 150 rpm
awRR Accuracy:	
0 to 70 rpm:	±1 rpm
71 to 120 rpm:	±2 rpm
121 to 150 rpm:	±3 rpm
Alarm:	10, 15, 20, 25, 30, 35, 40 s

Anesthesia Gases (ePM 12M and ePM 15M only)

Sampling Rate

Adult/Pediatric:	200 ml/min
Neonate:	120 ml/min
Sampling Rate Tolerance:	±10 ml/min or ±10%, whichever is greater
Delay Time:	<4 s
Warm-up Time:	45 s to iso accuracy mode
	10 min to full accuracy mode

Measurement Range

CO_2:	0 to 30%
N_2O:	0 to 100%
Des/Sev/Enf/Iso/Hal:	0 to 30%
O_2:	0 to 100%
awRR:	2 to 100 rpm

mindray NORTH AMERICA

ePM 10M/12M/15M PATIENT MONITORS

Performance Specifications for Ambulatory Surgery Center and Clinic

Anesthesia Gases (continued)

Resolution

CO_2:	0.1%
N_2O:	1%
Des/Sev/Enf/Iso/Hal:	0.1%
O_2:	1%
awRR:	1 rpm

Full Accuracy

Gases:	Range (% REL)	Accuracy (% ABS)
CO_2:	0 to 1%	±0.1%
	1 to 5%	±0.2%
	5 to 7%	±0.3%
	7 to 10%	±0.5%
	>10%	Not specified
N_2O:	0 to 20%	±2%
	20 to 100%	±3%
Des:	0 to 1%	±0.15%
	1 to 5%	±0.2%
	5 to 10%	±0.4%
	10 to 15%	±0.6%
	15 to 18%	±1%
	>18%	Not specified
Sev:	0 to 1%	±0.15%
	1 to 5%	±0.2%
	5 to 8%	±0.4%
	>8%	Not specified
Enf/Iso/Hal:	0 to 1%	±0.15%
	1 to 5%	±0.2%
	>5%	Not specified
O_2:	0 to 25%	±1%
	25 to 80%	±2%
	80 to 100%	±3%
awRR:	2 to 60 rpm	±1 rpm
	>60 rpm	Not specified

Rise Time

Sampling flow 120 ml/min, using the Neonatal DRYLINE II™ watertrap and a Neonatal 2.5m sampling line:

CO_2/ N_2O:	≤250 ms
Iso/Hal/Sev/Des:	≤300 ms
Enf:	≤350 ms
O_2:	≤600 ms

Sampling flow 200ml/min, using the Adult/Pediatric DRYLINE II™ watertrap and an Adult 2.5m sampling line:

CO_2/ N_2O:	≤250 ms
Iso/Hal/Sev/Des:	≤300 ms
Enf:	≤350 ms
O_2:	≤500 ms

Sampling Delay Time

Sampling flow 120 ml/min, using the Neonatal DRYLINE II™ watertrap and a Neonatal 2.5m sampling line:

CO_2:	≤ 4 s
N_2O:	≤4.2 s
O_2:	≤4 s
Enf /Iso/Hal/Sev/Des:	≤4.4 s

Sampling flow 200ml/min, using the Adult/Pediatric DRYLINE II™ watertrap and an Adult 2.5m sampling line:

CO_2:	≤4.2 s
N_2O:	≤4.3 s
O_2:	≤4 s
Enf/Iso/Hal/Sev/Des:	≤4.5 s
Apnea Alarm Delay:	10, 15, 20, 25, 30, 35, 40 s

Data Review Storage

Trends Data:	Up to 120 hours @ 1 min
Events:	Up to 1000 events, including parameter alarms, arrhythmia events, technical alarms
NIBP:	Up to 1000 sets
Full Disclosure:	48 hours for all parameter waveforms
OxyCRG:	400 OxyCRG events
ST Review:	Up to 120 hours @ 1 min

Wi-Fi Communications

Protocol:	IEEE 802.11a/b/g/n
Modulation mode	
	DSSS and OFDM
Operating Frequency:	IEEE 802.11b/g/n (2.4G):
	ETSI/FCC/KC: 2.4 to 2.483 GHz
	MIC: 2.4 to 2.495 GHz
	IEEE 802.11a/n (5G):
	ETSI: 5.15 to 5.35 GHz, 5.47 to 5.725 GHz
	FCC: 5.15 to 5.35 GHz, 5.725 to 5.82 GHz
	MIC: 5.15 to 5.35 GHz
	KC: 5.15 to 5.35 GHz, 5.47 to 5.725 GHz, 5.725 to 5.82 GHz
Channel Spacing:	5 MHz @ 2.4 GHz, 20 MHz @ 5 GHz
Wireless Baud Rate:	IEEE 802.11a: 6 to 54 Mbps
	IEEE 802.11b: 1 to 11 Mbps
	IEEE 802.11g: 6 to 54 Mbps
	IEEE 802.11n: 6.5 to 72.2 Mbps
Output Power:	< 20 dBm (CE requirement: detection mode - RMS) < 30 dBm (FCC requirement: detection mode - peak power)
Operating Mode:	Infrastructure
Data Security:	WPA-PSK, WPA2-PSK, WPA-Enterprise, WPA2-Enterprise (EAP-FAST. EAP-TLS, EAP- TTLS, PEAP-GTC, PEAP-MSCHAPv2, PEAP-TLS, LEAP)
Encryption:	AES

Interfacing

Main Unit:	AC power connector (1)
	VGA port (1)
	USB 2.0 connector (2)
	Analog output/nurse call/defib. sync/serial port (1) Integrated module rack (1), for 2 slots
Barcode Scanner:	Supports 1D & 2D barcode
Thermal Recorder:	3 traces (paper 50 mm width, 20 m length)
Network Printer:	Support

Power Requirements

Line Voltage:	100 to 240 VAC (±10 %)
Maximum Current:	2.0 A
Frequency:	50/60 Hz (±3 Hz)
Battery:	Rechargeable lithium-ion battery, 4500mAh
	≥4 hours run time
Recharge Time (power off):	5 hours to 90%

mindray NORTH AMERICA

ePM 10M/12M/15M PATIENT MONITORS

Performance Specifications for Ambulatory Surgery Center and Clinic

Physical Specifications

Weight:

ePM 10M:	8.8 lbs (4.0 Kg)
ePM 12M:	10.6 lbs (4.8 Kg)
ePM 15M:	13.7 lbs (6.2 Kg)

(Standard configuration, excluding modules, recorder, battery and accessories.)

Size:

ePM 10M:	10.6" x 9.9" x 6.3" 26.9 cm x 25.2 cm x 15.9 cm
ePM 12M:	12.2" x 11.4" x 6.7" 31.0 cm x 28.9 cm x 16.9 cm
ePM 15M:	15.5" x 12.4" x 6.8" 39.4 cm × 31.6 cm × 17.4 cm

Environmental Requirements

Temperature

Operating:	0 to 40 °C (without AG and CO_2 module)
	10 to 40 °C (with AG and CO_2 module)
Storage:	-20 to 60 °C

Humidity

Operating:	15 to 95 % (non condensing)
Storage:	10 to 95 % (non condensing)

Barometric

Operating:	427.5 to 805.5 mmHg (57.0 to 107.4 kPa, without AG and CO_2 module)
Storage:	120 to 805.5 mmHg (16.0 to 107.4 kPa, without AG and CO_2 module)

Mindray DS USA, Inc.
800 MacArthur Blvd., Mahwah, NJ 07430
Tel: 800.288.2121 Tel: 201.995.8000 Fax: 800.926.4275 **www.mindray.com**

©2024 Mindray DS USA, Inc.
Mindray® is a registered trademark of Shenzhen Mindray Bio-Medical Electronics Co., Ltd. All rights reserved.
All brands and product names are trademarks of their respective owners. Specifications subject to change without notice.
Contact your Mindray Representative for the most current information. 05/24 P/N: 0002-08-8076 Rev 1.0